GROUPE SEB - CALOR
ZI Monplaisir,
Rue du champ de courses
38780 Pont Evêque
FRANCE

Report N° 1174798A01 v3

EFFICACITE VIRUCIDE SUR DEFROISSEURS DR8135D1

30 September 2020

This report supersedes report 1174798A01 version n°2 which has to be destroyed. The laboratory absolves itself from the use of any previous report.

For the attention of **Cédric METAY**
GROUPE SEB - CALOR

Quotation 2020/62562 (DSP 767540)
Reference: TESTS ANTI VRUS COMMANDE N° 9324053626 09.06.2020

Tested product

Designation: DEFROISSEUR ACCESS TEAM DR8135D1
Reference: -
Batch N°: -
Brand: -
ATS reference: 776834

Study followed by Michel LEBREUIL

The copy of this report is only authorized by unabridged edition
This edition includes 14 pages

Eurofins ATS
505 rue Louis Berton
CS 50550
13594 AIX-EN-PROVENCE Cedex 3 – FRANCE
Tel : +33 (0)4 42 39 78 08
N° SIRET : 33761796300117

S.A.S AU CAPITAL DE 571 200 euros
Code APE : 7120B
ATS@eurofins.com

Eurofins ATS site de microbiologie
Bâtiment Cèdre 1
97 allée Alexandre Borodine
69800 SAINT PRIEST- FRANCE
Tel : +33 (0)4 26 78 48 88
N° SIRET : 33761796300133

eurofins | BioPharma Product Testing

SPONSOR	Eurofins ATS France		
	Aix-en-Provence, Pole d'activities Aix-Les-Milles		
	Actimart -1140 Rue Ampere		
	F-13851 Aix –en-Provence Cedex 3		
	FRANCE		
TEST METHOD	Antiviral activity on DEFROISSEUR ACCES STEAM DR8135D1 according to Sponsor's indications		
TEST ITEM			
PRODUCT NAME (*)	*DEFROISSEUR ACCES STEAM DR8135D1*		
MATRIX OF THE PRODUCT (*)	N.A.		
BATCH (*)	N.A.	ATS CODE	776834
MANUFACTURING DATE (*)	N.A.	EXPIRY DATE (*)	N.A.
MANUFACTURER	ROWENTA		
ACTIVE INGREDIENT (*)	N.A.		
MATERIAL ITEM ALIQUOT	LV-MAT-IJE2-20-204-0F90:a		
PARCEL REGISTRATION N.	IP-LV-2020199-AMD	RECEIVING DATE	17-July-2020
STORAGE CONDITIONS (*)	Room temperature (20°C± 5°C)		
(*) INFORMATION PROVIDED BY THE SPONSOR			
ANALYSIS STARTING DATE	18-August-2020	**ANALYSIS ENDING DATE**	26-August-2020
EXPERIMENTAL CONDITIONS			
SUMMARY	This study has been conducted in order to evaluate the antiviral activity (expressed in LogTCID$_{50}$ reduction value) of DEFROISSEUR ACCES STEAM DR8135D1 on contaminated fabric, compared to virus control (recovery after drying), following the Sponsor's ironing procedure and according to Sponsor's requirements. For this purpose, the following viruses have been used: *Murine norovirus (MNV, strain S99)* RVB-651 *Betacoronavirus 1 (Bovine Corona Virus) strain S379 Riems* RVB-0020 For each virus, 9 unfinished 100% cotton swatches (2x5 cm each) have been inoculated with 0,05 ml of viral suspension (as per EN 16777:2018 / UNI EN 16777:2019).		

eurofins | BioPharma Product Testing

	- 3 swatches have been inoculated and immediately the viable viral particles have been recovered in order to verify the recovery method from swatches. - 3 swatches have been inoculated and left in an opened petri dish on surface of a Biohazard hood until visible dry; then, immediately after drying, the viable viral particles have been recovered in order to verify the recovery after drying from swatches. - 3 swatches have been inoculated and left in an opened petri dish on surface of a Biohazard hood until visible dry. Then, they have been put in vertical position, treated with 4 applications of continuous steam (2 forwards and 2 backwards), spaced by 1 second (without steam), at 1 cm/sec remaining in contact with the cotton swatches, according to the Sponsor requirements. Waiting time between each carrier: 1 minute without steam. - The steam generating device was set to “Delicate”, that is the setting when the appliance is turned on.		
TEST TEMPERATURE	“Delicate” on the device (as per Sponsor instructions)		
PRODUCT APPEARANCE	Steam iron		
EXPOSURE SPEED	1 cm/sec		
CARRIERS	2 cm × 5 cm cotton fabric (swatches)		
VIRUS RECOVERY	Iced culture Medium		
INTERFERING SUBSTANCE	No interfering substance		
INCUBATION TEMPERATURE	37°C ± 1°C (with 5% CO_2)		
TEST VIRUSES	*Murine norovirus (MNV, strain S99), RVB-651* *Betacoronavirus 1 (Bovine Corona Virus) strain S379 Riems, RVB-0020*		
CELL LINES	*RAW 264.7, ATCC TIB-71* *PT, CCLV-RIE 11*		
		Log reductions	**% reduction**
RESULTS	*Murine norovirus, MNV strain S99*	**2.62 ± 0.104**	**99.76%**
	Betacoronavirus 1 (Bovine Corona Virus) strain S379 Riems	**≥3.06 ± 0.040**	**>99.9%**
	See Addendum N.1		
CONCLUSIONS	**DEFROISSEUR ACCES STEAM DR8135D1 CAUSES a reduction of 2.62 ± 0.104 Log (99.76%)** against *Murine norovirus, MNV strain S99 RVB-651* with 4 applications of continuous steam (2 forwards and 2 backwards), spaced by 1 second (without steam), at 1 cm/sec remaining in contact with the cotton swatches, compared to virus control (recovery after drying). Nevertheless, some residual virus is detected. **DEFROISSEUR ACCES STEAM DR8135D1 CAUSES a reduction ≥3.06 ± 0.040 Log (>99.9%)** against *Betacoronavirus 1 (Bovine Corona Virus) strain S379 Riems RVB-0020* with 4 applications of continuous steam (2 forwards and 2 backwards), spaced by 1 second (without steam), at 1 cm/sec remaining in contact with the cotton swatches, compared to virus control (recovery after drying).		

	The vapor temperature on the treated surface and the net exposure time are typically the critical parameters of the process. The higher these values, the greater the effectiveness of the process. Indeed, *Murine norovirus* is a non-enveloped virus, which makes it more resistant to disinfection than *Betacoronavirus 1*, which is an enveloped virus. The envelope is a structure that has the main function of facilitating the infection of the virus to the host, but which nevertheless makes the virus more sensitive to chemical agents.
ADDENDA	N. 1: RAW DATA ELABORATION (*5 PAGES*) ON BETACORONAVIRUS N. 2: RAW DATA ELABORATION (*5 PAGES*) ON MURINE NOROVIRUS

Reviewed and electronically signed for Study Technical Supervisor Approval by
Andrea De Bortoli, Employee
for Eurofins Biolab Srl, on 29-Sep-2020 15:54:26 UTC+02:00

eurofins

Quantitative non-porous surface test without mechanical action for the evaluation of virucidal activity of chemical disinfectants used in the medical area — Test method and requirements (phase2/step2)

Norma (Standard): EN16777:2018/ UNI EN16777:2019

Data inizio *(Started on)*: 19/08/20

Data fine test (*Test finished on*): 24/08/20

Rapporto No *(Report No)*: STULV20AA3931-1

ID Campione *(ID sample)*: LV-MAT-IJE2-20-204-0F90:a

Titolazione virus *(Virus Titration)*
Betacoronavirus 1 (Bovine Corona Virus) strain S379 Riems RVB-0020

Condizioni testate *(Test condition)*	Replica	K-	Diluizione virus *(Virus dilution)*								K-
			1	2	3	4	5	6	7	8	
Betacoronavirus 1 (Bovine Corona Virus) strain S379 Riems RVB-0020	**B**	0	4	4	4	4	4	4	0	0	0
	C	0	4	4	4	4	4	0	0	0	0
	D	0	4	4	4	4	4	0	0	0	0
	E	0	4	4	4	4	4	4	0	0	0
	F	0	4	4	4	4	4	4	0	0	0
	G	0	4	4	4	4	4	0	0	0	0
	Endpoint	0.0	100.0	100.0	100.0	100.0	100.0	50.0	0.0	0.0	0.0

Cell destruction: **VALID**

Log TCID50: **6.00** ± **0.447**

Data verifica Approver (*Approver verification date*): 26/08/20

Sigla tecnico *(Technician signature)*: _____ Data *(Date)*: 04/09/20

Sigla Approver *(Approver signature)*: _____ Data *(Date)*: 04/09/20

eurofins	Quantitative non-porous surface test without mechanical action for the evaluation of virucidal activity of chemical disinfectants used in the medical area — Test method and requirements (phase2/step2) Norma (Standard): EN16777:2018/ UNI EN16777:2019	Pagina *(Page)* 2 / 5

Data inizio *(Started on)*: 19/08/20 Data fine test (*Test finished on*): 24/08/20

Rapporto No *(Report No)*: STULV20AA3931-1 ID Campione *(ID sample)*: LV-MAT-IJE2-20-204-0F90:a

Recovery after inoculum from swatches
Dilution in ice-cold medium
Betacoronavirus 1 (Bovine Corona Virus) strain S379 Riems RVB-0020

Carrier 1
Condizioni testate *(Test condition)*

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	4	4	0	0	0	0	0
C	0	4	4	4	3	0	0	0	0	0
D	0	4	4	4	4	0	0	0	0	0
E	0	4	4	4	3	0	0	0	0	0
F	0	4	4	4	4	0	0	0	0	0
G	0	4	4	4	3	0	0	0	0	0
Endpoint	0.0	100.0	100.0	100.0	100.0	0.0	0.0	0.0	0.0	0.0

No interfering substance
0 min

Cell destruction: **VALID**
Log TCID50: **5.80** ± **0.000**

Carrier 2
Condizioni testate *(Test condition)*

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	4	4	0	0	0	0	0
C	0	4	4	4	3	0	0	0	0	0
D	0	4	4	4	0	0	0	0	0	0
E	0	4	4	4	3	0	0	0	0	0
F	0	4	4	4	3	0	0	0	0	0
G	0	4	4	4	4	0	0	0	0	0
Endpoint	0.0	100.0	100.0	100.0	83.3	0.0	0.0	0.0	0.0	0.0

No interfering substance
0 min

Cell destruction: **VALID**
Log TCID50: **5.63** ± **0.346**

Carrier 1
Condizioni testate *(Test condition)*

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	4	4	3	0	0	0	0
C	0	4	4	4	3	0	0	0	0	0
D	0	4	4	4	4	0	0	0	0	0
E	0	4	4	4	4	0	0	0	0	0
F	0	4	4	4	4	0	0	0	0	0
G	0	4	4	4	3	0	0	0	0	0
Endpoint	0.0	100.0	100.0	100.0	100.0	16.7	0.0	0.0	0.0	0.0

No interfering substance
0 min

Cell destruction: **VALID**
Log TCID50: **5.97** ± **0.346**
Log TCID50 (Average): **5.80** ± **0.200**

Data verifica Approver (*Approver verification date*): 26/08/20

Sigla tecnico *(Technician signature)*: _____ Data *(Date)*: 04/09/20

Sigla Approver *(Approver signature)*: _____ Data *(Date)*: 04/09/20

eurofins

Quantitative non-porous surface test without mechanical action for the evaluation of virucidal activity of chemical disinfectants used in the medical area — Test method and requirements (phase2/step2)

Norma (Standard): EN16777:2018/ UNI EN16777:2019

Data inizio *(Started on)*: 19/08/20 Data fine test (*Test finished on*):

Rapporto No *(Report No)*: STULV20AA3931-1 ID Campione *(ID sample)*: LV-MAT-IJE2-20-204-0F90:a

Recovery after drying from swatches
Dilution in ice-cold medium
Betacoronavirus 1 (Bovine Corona Virus) strain S379 Riems RVB-0020

Carrier 1
Condizioni testate *(Test condition)*

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	4	0	0	0	0	0	0
C	0	4	4	3	0	0	0	0	0	0
D	0	4	4	4	0	0	0	0	0	0
E	0	4	4	4	0	0	0	0	0	0
F	0	4	4	4	0	0	0	0	0	0
G	0	4	4	3	0	0	0	0	0	0
Endpoint	0.0	100.0	100.0	100.0	0.0	0.0	0.0	0.0	0.0	0.0

No interfering substance

Cell destruction: VALID
Log TCID50: **4.80** ± **0.000**

Carrier 2
Condizioni testate *(Test condition)*

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	4	0	0	0	0	0	0
C	0	4	4	3	0	0	0	0	0	0
D	0	4	4	4	0	0	0	0	0	0
E	0	4	4	4	0	0	0	0	0	0
F	0	4	4	4	3	0	0	0	0	0
G	0	4	4	4	0	0	0	0	0	0
Endpoint	0.0	100.0	100.0	100.0	16.7	0.0	0.0	0.0	0.0	0.0

No interfering substance

Cell destruction: VALID
Log TCID50: **4.97** ± **0.346**

Carrier 1
Condizioni testate *(Test condition)*

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	3	0	0	0	0	0	0
C	0	4	4	3	0	0	0	0	0	0
D	0	4	4	4	0	0	0	0	0	0
E	0	4	4	4	0	0	0	0	0	0
F	0	4	4	4	0	0	0	0	0	0
G	0	4	4	4	0	0	0	0	0	0
Endpoint	0.0	100.0	100.0	100.0	0.0	0.0	0.0	0.0	0.0	0.0

No interfering substance

Cell destruction: VALID
Log TCID50: **4.80** ± **0.000**
Log TCID50 (Average): **4.86** ± **0.141**

Data verifica Approver (*Approver verification date*): 26/08/20

Sigla tecnico *(Technician signature)*: _____ Data *(Date)*: 26.08.20

Sigla Approver *(Approver signature)*: _____ Data *(Date)*: 26/08/20

eurofins

Quantitative non-porous surface test without mechanical action for the evaluation of virucidal activity of chemical disinfectants used in the medical area — Test method and requirements (phase2/step2)

Norma (Standard): EN16777:2018/ UNI EN16777:2019

Data inizio *(Started on)*: 19/08/20 Data fine test (*Test finished on*): 24/08/20

Rapporto No *(Report No)*: STULV20AA3931-1 ID Campione *(ID sample)*: LV-MAT-IJE2-20-204-0F90:a

Procedura test *(Test procedure)*
Dilution in ice-cold medium
Betacoronavirus 1 (Bovine Corona Virus) strain S379 Riems RVB-0020

Carrier 1
Condizioni testate *(Test condition)*

DEFROISSEUR ACCES STEAM DR8135D1
N.A.
No interfering substance
1 cm/sec

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	0	0	0	0	0	0	0	0	0
C	0	0	0	0	0	0	0	0	0	0
D	0	0	0	0	0	0	0	0	0	0
E	0	0	0	0	0	0	0	0	0	0
F	0	0	0	0	0	0	0	0	0	0
G	0	0	0	0	0	0	0	0	0	0
Endpoint	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0

Cell destruction: **VALID**
Log TCID50: ≤ **1.80** ± **0.000**

Carrier 2
Condizioni testate *(Test condition)*

DEFROISSEUR ACCES STEAM DR8135D1
N.A.
No interfering substance
1 cm/sec

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	0	0	0	0	0	0	0	0	0
C	0	0	0	0	0	0	0	0	0	0
D	0	0	0	0	0	0	0	0	0	0
E	0	0	0	0	0	0	0	0	0	0
F	0	0	0	0	0	0	0	0	0	0
G	0	0	0	0	0	0	0	0	0	0
Endpoint	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0

Cell destruction: **VALID**
Log TCID50: ≤ **1.80** ± **0.000**

Carrier 3
Condizioni testate *(Test condition)*

DEFROISSEUR ACCES STEAM DR8135D1
N.A.
No interfering substance
1 cm/sec

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	0	0	0	0	0	0	0	0	0
C	0	0	0	0	0	0	0	0	0	0
D	0	0	0	0	0	0	0	0	0	0
E	0	0	0	0	0	0	0	0	0	0
F	0	0	0	0	0	0	0	0	0	0
G	0	0	0	0	0	0	0	0	0	0
Endpoint	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0

Cell destruction: **VALID**
Log TCID50: ≤ **1.80** ± **0.000**
Log TCID50 (Average): ≤ **1.80** ± **0.000**
Reduction (Average): ≥ **3.06** ± **0.040**

Data verifica Approver (*Approver verification date*): 26/08/20

Sigla tecnico *(Technician signature)*: _____ Data *(Date)*: 04/09/20

Sigla Approver *(Approver signature)*: _____ Data *(Date)*: 04/09/20

eurofins	Quantitative non-porous surface test without mechanical action for the evaluation of virucidal activity of chemical disinfectants used in the medical area — Test method and requirements (phase2/step2) Norma (Standard): EN16777:2018/ UNI EN16777:2019	Pagina *(Page)* 5 / 5

Data inizio *(Started on)*: 19/08/20 Data fine test (*Test finished on*): 24/08/20

Rapporto No *(Report No)*: STULV20AA3931-1 ID Campione *(ID sample)*: LV-MAT-IJE2-20-204-0F90:a

Result summary

Attività virucida *(Virucidal activity)*
Betacoronavirus 1 (Bovine Corona Virus) strain S379 Riems RVB-0020

Prodotto *(Product)*	**DEFROISSEUR ACCES STEAM DR8135D1**
Sostanza interferente *(Interfering substance)*	**No interfering substance**
Velocità *(speed)*	1 cm/sec
Concentrazione *(Concentration)*	**Riduzione Log *(Log Reduction)***
N.A.	≥ 3.06 ± 0.04

≥	3.06		±	0.04

Data verifica Approver (*Approver verification date*): 26/08/20

Sigla Approver *(Approver signature)*: A Data *(Date)*: 04/09/20

eurofins

Quantitative non-porous surface test without mechanical action for the evaluation of virucidal activity of chemical disinfectants used in the medical area — Test method and requirements (phase2/step2)

Norma (Standard): EN16777:2018/ UNI EN16777:2019

Data inizio *(Started on)*: 21/08/20

Data fine test (*Test finished on*): 24/08/20

Rapporto No *(Report No)*: STULV20AA3931-1

ID Campione *(ID sample)*: LV-MAT-IJE2-20-204-0F90:a

Titolazione virus *(Virus Titration)*
Murine norovirus (MNV, strain S99) RVB-651

Condizioni testate *(Test condition)*	Replica	K-	Diluizione virus *(Virus dilution)*								K-
			1	2	3	4	5	6	7	8	
Murine norovirus (MNV, strain S99) RVB-651	B	0	4	4	4	4	4	0	0	0	0
	C	0	4	4	4	4	4	4	0	0	0
	D	0	4	4	4	4	4	4	0	0	0
	E	0	4	4	4	4	4	4	0	0	0
	F	0	4	4	4	4	4	4	0	0	0
	G	0	4	4	4	4	4	4	0	0	0
	Endpoint	0.0	100.0	100.0	100.0	100.0	100.0	83.3	0.0	0.0	0.0

Cell destruction: **VALID**

Log TCID50: **6.33** ± **0.346**

Data verifica Approver (*Approver verification date*): 26/08/20

Sigla tecnico *(Technician signature)*: _____ Data *(Date)*: 26/08/20

Sigla Approver *(Approver signature)*: _____ Data *(Date)*: 26/08/20

eurofins

Quantitative non-porous surface test without mechanical action for the evaluation of virucidal activity of chemical disinfectants used in the medical area — Test method and requirements (phase2/step2)

Norma (Standard): EN16777:2018/ UNI EN16777:2019

Data inizio *(Started on)*: 21/08/20

Data fine test (*Test finished on*): 24/08/20

Rapporto No *(Report No)*: STULV20AA3931-1

ID Campione *(ID sample)*: LV-MAT-IJE2-20-204-0F90:a

Recovery after inoculum from swatches
Dilution in ice-cold medium
Murine norovirus (MNV, strain S99) RVB-651

Carrier 1
Condizioni testate *(Test condition)*

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	4	4	0	0	0	0	0
C	0	4	4	4	4	3	0	0	0	0
D	0	4	4	4	4	3	0	0	0	0
E	0	4	4	4	4	0	0	0	0	0
F	0	4	4	4	4	0	0	0	0	0
G	0	4	4	4	4	0	0	0	0	0
Endpoint	0.0	100.0	100.0	100.0	100.0	33.3	0.0	0.0	0.0	0.0

No interfering substance
0 min

Cell destruction: **VALID**
Log TCID50: **6.13** ± **0.400**

Carrier 2
Condizioni testate *(Test condition)*

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	4	4	0	0	0	0	0
C	0	4	4	4	4	0	0	0	0	0
D	0	4	4	4	4	0	0	0	0	0
E	0	4	4	4	4	4	0	0	0	0
F	0	4	4	4	4	4	0	0	0	0
G	0	4	4	4	4	0	0	0	0	0
Endpoint	0.0	100.0	100.0	100.0	100.0	33.3	0.0	0.0	0.0	0.0

No interfering substance
0 min

Cell destruction: **VALID**
Log TCID50: **6.13** ± **0.400**

Carrier 1
Condizioni testate *(Test condition)*

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	4	4	4	0	0	0	0
C	0	4	4	4	4	4	0	0	0	0
D	0	4	4	4	4	0	0	0	0	0
E	0	4	4	4	4	0	0	0	0	0
F	0	4	4	4	4	4	0	0	0	0
G	0	4	4	4	4	4	0	0	0	0
Endpoint	0.0	100.0	100.0	100.0	100.0	66.7	0.0	0.0	0.0	0.0

No interfering substance
0 min

Cell destruction: **VALID**
Log TCID50: **6.47** ± **0.400**
Log TCID50 (Average): **6.24** ± **0.283**

Data verifica Approver (*Approver verification date*): 26/08/20

Sigla tecnico *(Technician signature)*: _____ Data *(Date)*: 04.09.20

Sigla Approver *(Approver signature)*: _____ Data *(Date)*: 04/09/20

eurofins

Quantitative non-porous surface test without mechanical action for the evaluation of virucidal activity of chemical disinfectants used in the medical area — Test method and requirements (phase2/step2)

Norma (Standard): EN16777:2018/ UNI EN16777:2019

Data inizio *(Started on)*: 21/08/20

Data fine test (*Test finished on*):

Rapporto No *(Report No)*: STULV20AA3931-1

ID Campione *(ID sample)*: LV-MAT-IJE2-20-204-0F90:a

Recovery after drying from swatches
Dilution in ice-cold medium
Murine norovirus (MNV, strain S99) RVB-651

Carrier 1
Condizioni testate *(Test condition)*

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	4	4	0	0	0	0	0
C	0	4	4	4	4	4	0	0	0	0
D	0	4	4	4	4	0	0	0	0	0
E	0	4	4	4	4	0	0	0	0	0
F	0	4	4	4	4	4	0	0	0	0
G	0	4	4	4	4	0	0	0	0	0
Endpoint	0.0	100.0	100.0	100.0	100.0	33.3	0.0	0.0	0.0	0.0

No interfering substance

Cell destruction: **VALID**
Log TCID50: **6.13** ± **0.400**

Carrier 2
Condizioni testate *(Test condition)*

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	4	4	0	0	0	0	0
C	0	4	4	4	4	0	0	0	0	0
D	0	4	4	4	4	3	0	0	0	0
E	0	4	4	4	4	0	0	0	0	0
F	0	4	4	4	4	4	0	0	0	0
G	0	4	4	4	4	0	0	0	0	0
Endpoint	0.0	100.0	100.0	100.0	100.0	33.3	0.0	0.0	0.0	0.0

No interfering substance

Cell destruction: **VALID**
Log TCID50: **6.13** ± **0.400**

Carrier 1
Condizioni testate *(Test condition)*

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	4	4	0	0	0	0	0
C	0	4	4	4	4	3	0	0	0	0
D	0	4	4	4	4	0	0	0	0	0
E	0	4	4	4	4	0	0	0	0	0
F	0	4	4	4	4	0	0	0	0	0
G	0	4	4	4	4	0	0	0	0	0
Endpoint	0.0	100.0	100.0	100.0	100.0	16.7	0.0	0.0	0.0	0.0

No interfering substance

Cell destruction: **VALID**
Log TCID50: **5.97** ± **0.346**
Log TCID50 (Average): **6.08** ± **0.271**

Data verifica Approver (*Approver verification date*): 26/08/20

Sigla tecnico *(Technician signature)*: _____ Data *(Date)*: 06/09/20

Sigla Approver *(Approver signature)*: _____ Data *(Date)*: 06/09/20

eurofins

Quantitative non-porous surface test without mechanical action for the evaluation of virucidal activity of chemical disinfectants used in the medical area — Test method and requirements (phase2/step2)

Norma (Standard): EN16777:2018/ UNI EN16777:2019

Data inizio *(Started on)*: 21/08/20

Data fine test (*Test finished on*): 24/08/20

Rapporto No *(Report No)*: STULV20AA3931-1

ID Campione *(ID sample)*: LV-MAT-IJE2-20-204-0F90:a

Procedura test *(Test procedure)*
Dilution in ice-cold medium
Murine norovirus (MNV, strain S99) RVB-651

Carrier 1
Condizioni testate *(Test condition)*

DEFROISSEUR ACCES STEAM DR8135D1
N.A.
No interfering substance
1 cm/sec

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	0	0	0	0	0	0	0
C	0	4	0	0	0	0	0	0	0	0
D	0	4	0	0	0	0	0	0	0	0
E	0	4	0	0	0	0	0	0	0	0
F	0	4	4	0	0	0	0	0	0	0
G	0	4	0	0	0	0	0	0	0	0
Endpoint	0.0	100.0	33.3	0.0	0.0	0.0	0.0	0.0	0.0	0.0

Cell destruction: **VALID**
Log TCID50: **3.13** ± **0.400**

Carrier 2
Condizioni testate *(Test condition)*

DEFROISSEUR ACCES STEAM DR8135D1
N.A.
No interfering substance
1 cm/sec

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	0	0	0	0	0	0	0	0
C	0	4	4	0	0	0	0	0	0	0
D	0	4	4	0	0	0	0	0	0	0
E	0	4	4	0	0	0	0	0	0	0
F	0	4	4	0	0	0	0	0	0	0
G	0	4	4	0	0	0	0	0	0	0
Endpoint	0.0	100.0	83.3	0.0	0.0	0.0	0.0	0.0	0.0	0.0

Cell destruction: **VALID**
Log TCID50: **3.63** ± **0.346**

Carrier 3
Condizioni testate *(Test condition)*

DEFROISSEUR ACCES STEAM DR8135D1
N.A.
No interfering substance
1 cm/sec

Replica	K-	Diluizione virus (Virus dilution)								K-
		2.3	3.3	4.3	5.3	6.3	7.3	8.3	9.3	
B	0	4	4	0	0	0	0	0	0	0
C	0	4	4	0	0	0	0	0	0	0
D	0	4	4	0	0	0	0	0	0	0
E	0	4	4	0	0	0	0	0	0	0
F	0	4	0	0	0	0	0	0	0	0
G	0	4	4	0	0	0	0	0	0	0
Endpoint	0.0	100.0	83.3	0.0	0.0	0.0	0.0	0.0	0.0	0.0

Cell destruction: **VALID**
Log TCID50: **3.63** ± **0.346**
Log TCID50 (Average): **3.46** ± **0.258**
Reduction (Average): **2.62** ± **0.104**

Data verifica Approver (*Approver verification date*): 26/08/20

Sigla tecnico *(Technician signature)*: _____ Data *(Date)*: 04/09/20

Sigla Approver *(Approver signature)*: _____ Data *(Date)*: 04/09/20

eurofins	Quantitative non-porous surface test without mechanical action for the evaluation of virucidal activity of chemical disinfectants used in the medical area — Test method and requirements (phase2/step2) Norma (Standard): EN16777:2018/ UNI EN16777:2019	Pagina *(Page)* 5 / 5

Data inizio *(Started on)*: 21/08/20 Data fine test (*Test finished on*): 24/08/20

Rapporto No *(Report No)*: STULV20AA3931-1 ID Campione *(ID sample)*: LV-MAT-IJE2-20-204-0F90:a

Result summary

Attività virucida *(Virucidal activity)*
Murine norovirus (MNV, strain S99) RVB-651

Prodotto *(Product)*	**DEFROISSEUR ACCES STEAM DR8135D1**
Sostanza interferente *(Interfering substance)*	**No interfering substance**
Velocità *(speed)*	1 cm/sec
Concentrazione *(Concentration)*	**Riduzione Log *(Log Reduction)***
N.A.	2.62 ± 0.104

	2.62		±	0.10

Data verifica Approver (*Approver verification date*): 26/08/20

Sigla Approver *(Approver signature)*: A Data *(Date)*: 24/09/20